



Sponsored by Alcon Laboratories, Inc.

CENTURION Vision System: My Experience

Experience with torsional energy.

BY DANIELE TOGNETTO, MD

I have been using INFINITI Vision System with OZil technology since it was launched in 2006. From my first hours, I noticed a significant improvement in followability and a reduction in repulsion compared to my experience using longitudinal phacoemulsification.¹ At that time, I got the impression that this torsional tip movement was an innovation in cataract removal surgery. Over time, I have come to believe the torsional modality leads to a reduction in ultrasound time (UST) and cumulative dissipated energy (CDE) compared to longitudinal modality,¹⁻³ since the torsional tip movement can result in less chattering of material at the tip as the back-and-forward movement of longitudinal modality is replaced by the side-to-side tip movement of torsional modality. The reduction of chattering^{4,5} leads to an increase in removal efficiency of the cataract removal procedure.^{1,2,6}

The use of the torsional ultrasound modality has been associated in the literature with a reduced loss of endothelial cells and tissue damage compared to the traditional longitudinal modality's literature.^{1,2,6} Also, the thermal profile of the tip's movement has been shown to be lower using the torsional modality over the longitudinal modality.^{7,8}

In 2012, colleagues and I published in the *Journal of Cataract & Refractive Surgery* a study entitled, "Stroke Dynamics and Frequency of 3 Phacoemulsification Machines,"⁹ in which we compared the working frequency and the stroke dynamics of the phaco tip for three different systems. In contrast to other phaco delivery systems, The INFINITI Vision System with OZil technology showed a progressively increasing lateral movement according to the increased phaco power, almost without any longitudinal movement, both at the end of the tip and at the incision site.

The torsional tip's movement may provoke a kind of "fogging" effect due to the pulverization of the lens material, reducing visibility in the anterior chamber. Also, high-vacuum procedures using flared or tapered tips may lead to material not moving completely through the tip. This occlusion by small, hard lens fragments could make it necessary to stop the procedure and clear the obstruction. These effects were overcome in 2009 with the innovative OZil IP (Intelligent Phaco) software (Alcon). This enhancement helps surgeons avoid strong occlusions by using a longitudinal pulse of energy to produce a burst of repulsion when a preset vacuum threshold is reached.^{10,11}

In my clinic's operating theater, we use different phaco machines. I exclusively use INFINITI Vision System with OZil IP. I was impressed by the efficiency and performance of this machine; so much so, I did not think that we could get any better.

MY EXPERIENCE WITH CENTURION VISION SYSTEM

In 2013, Alcon launched the CENTURION Vision System. From my early experience with the CENTURION Vision System, I feel it provides two main technical innovations that make this system unique and revolutionary:

- The Active Fluidics Technology
- The CENTURION Energy Delivery, especially utilizing the INTREPID BALANCED Tip

ACTIVE FLUIDICS TECHNOLOGY

In the CENTURION Vision System, the Active Fluidics technology means that the irrigation inflow is not passively linked to the static height of the bottle, but is controlled by dynamic compression of a soft bag by a plate inside the system. This allows the CENTURION Vision System to rapidly detect and compensate for dynamic changes that affect the anterior chamber, leading to enhanced chamber stability.

The rapid compensation is governed by the information derived from a combination of a vacuum sensor placed along the aspiration line, the known flow rate, and by a pressure sensor placed along the irrigation line. This technology, which responds to very small changes in pressure, allows the surgeon to select the target pressure inside the eye throughout the procedure. The system also incorporates a unique low-compliance design that reduces the storage of potential energy and postocclusion surge. This allows two surgical advantages: enhanced control of intraocular pressure throughout all stages of the procedure, and the option to operate at a lower surgical target pressure.

CENTURION ENERGY DELIVERY

The second innovation of CENTURION Vision System is in energy delivery. I used the CENTURION Vision System with the INTREPID BALANCED tip, an innovative design that has been specifically designed to enhance the characteristics of the OZil torsional motion. The INTREPID BALANCED Tip has two opposite bends along its axis, with an overall profile similar to

Advanced-Technology Cataract Surgery

CENTURION VISION SYSTEM ENERGY DELIVERY

CENTURION Energy Delivery combines existing and new capabilities to enhance the efficiency of cataract removal:

1. The Active Fluidics capabilities of CENTURION Vision System.
2. OZil torsional technology (including OZil IP).
3. And especially with the enhanced performance of the BALANCED tip.

With the CENTURION Vision System, in addition to the Active Fluidics technology designed to maintain chamber stability, the surgeon can choose CENTURION Energy Delivery capabilities in aspiration and vacuum to help

- bring materials to the tip;
- have materials at the tip behave the way the surgeon prefers;
- move material through the tip.

With INFINITI Vision System, vacuum and aspiration can be linear (or fixed) in foot pedal position 2 (FP 2). In FP 3, the surgeon-selected parameters are always fixed.

With the CENTURION Vision System, vacuum and aspiration can be linear (or fixed) in FP 2 and, can be either:

- Fixed in FP 3: Like INFINITI Vision System
- Decreasing in FP 3: This ability, unique to CENTURION System, allows the surgeon to reduce vacuum and flow while applying

phaco energy. It is designed to reduce the surge at the occlusion break.

- Increasing in FP 3: This ability, unique to CENTURION System, allows surgeon to “grab” and “hold” fragments for repositioning and the next round of phaco and aspiration.

These choices help the surgeon achieve his or her preference regarding how to manage the interaction of bringing materials to the tip (via aspiration), holding material to allow desired repositioning (vacuum), along with emulsification and repulsion of material at the tip (phaco energy delivery), and finally removal from the anterior chamber (aspiration), with a wider range of choice than is available on INFINITI Vision System. The result is the desired use of cumulative dissipated energy (CDE) needed to achieve cataract removal efficiency.

These capabilities are added to the OZil IP (Intelligent Phaco), where a surgeon-selected amount of longitudinal energy can be employed to provide momentary repulsion when a vacuum threshold is achieved.

All the above work with all of the Alcon phaco tips. However, due to the design of the BALANCED tip, where the energy is focused at the distal end of the tip and the torsional amplitude is greater, the level of efficiency achieved is further enhanced.

a straight tip. The maximum lateral stroke of the tip is equal to 180 μm , significantly higher than the typical Kelman tip, whose stroke can reach 120 μm . The inner diameter of the tip is 0.57 mm, and the outer diameter is 0.8 mm, without any tapering or flaring along the length of the tip.

In my early experience with the CENTURION Vision System using this tip, I was impressed by its efficiency. The lenticular fragments seem to disappear once engaged by the tip. The rapidity of the removal of the lens goes hand-in-hand with the stability of the anterior chamber. The efficiency of the tip allows me to dissolve hard cataracts in a very short time without the use of high amounts of energy. I felt confident with a stable anterior chamber and had a sense of control throughout the procedure.

EXPERIENCE WITH INTREPID BALANCED TIP

In order to test the effectiveness of the CENTURION Vision System with the INTREPID BALANCED tip, we removed some hard cataracts. We first divided the nucleus into two parts, and then removed the first half using the Mini Flared Kelman Tip and the second half using the INTREPID BALANCED tip. The total phaco time was noticeably shorter using the INTREPID BALANCED tip,¹² which demonstrated the great efficiency of the new system and tip.

In my opinion, CENTURION Vision System advances the technological frontier in the field of cataract surgery.

Its use will be of a benefit for a wide range of surgical procedures for surgeons of all levels of experience. ■

Daniele Tognetto, MD, is Associate Professor of Ophthalmology at the Eye Clinic, University of Trieste, Ospedale Maggiore, Italy. He discloses no financial interest in the products or companies mentioned herein. Professor Tognetto may be reached at +39 407 72449; tognetto@univ.trieste.it.

1. Vasavada AR, Raj SM, Patel U. Comparison of torsional and microburst longitudinal phacoemulsification: a prospective, randomized, masked clinical trial. *Ophthalmic Surg Lasers Imaging*. 2010;41(1):109-114.
2. Liu Y, Zeng M, Liu X, et al. Torsional mode versus conventional ultrasound mode phacoemulsification: Randomized comparative clinical study. *J Cataract & Refract Surg*. 2007;33(2):287-292.
3. Rekas M, Montés-Micó R, Kirić-Jachym K. Comparison of torsional and longitudinal modes using phacoemulsification parameters. *J Cataract Refract Surg*. 2009;35(10):1719-1724.
4. Davison JA. Cumulative tip travel and implied followability of longitudinal and torsional phacoemulsification. *J Cataract Refract Surg*. 2008;34(6):986-990.
5. Fernández de Castro LE, Dimalanta RC, Solomon KD. Bead-flow pattern: Quantitation of fluid movement during torsional and longitudinal phacoemulsification. *J Cataract Refract Surg*. 2010;36(6):1018-1023.
6. Christakis PG, Braga-Mele RM. Intraoperative performance and postoperative outcome comparison of longitudinal, torsional, and transversal phacoemulsification machines. *J Cataract Refract Surg*. 2012;38(2):234-241.
7. Zacharias J. Thermal characterization of phaco probes operated in torsional and longitudinal modalities using alternative methodologies. Paper presented at: The annual symposium of the ASCRS, 2011; San Diego, CA.
8. Jun B, Berdahl JP, Kim T. Thermal study of longitudinal and torsional ultrasound phacoemulsification: tracking the temperature of the corneal surface, incision, and handpiece. *J Cataract Refract Surg*. 2010;36(5):832-837.
9. Tognetto D, Cecchini P, Leon P, et al. Stroke dynamics and frequency of 3 phacoemulsification machines. *J Cataract Refract Surg*. 2012;38(2):333-342.
10. Cionni R, A.S. Crandall, and D. Felsted. Length and frequency of intraoperative occlusive events with new torsional phacoemulsification software. *J Cataract Refract Surg*. 2011;37(10):1785-1790.
11. Ratnarajan G, Packard R, Ward M. Combined occlusion-triggered longitudinal and torsional phacoemulsification during coaxial microincision cataract surgery: Effect on 30-degree mini-flared tip behavior. *J Cataract Refract Surg*. 2011;37(5):825-829.
12. Solomon KD, Lorente R, Cionni RJ, Fanney D. Prospective, randomized clinical study using novel phaco system with intraocular system target pressure control and existing phaco system. Paper presented at: The ASCRS/ASOA Symposium & Congress, April 28, 2014; Boston.